

INDI: Instrument-Neutral Distributed Interface

Control Protocol, Version 1.2

Elwood Charles Downey

© 2003 Clear Sky Institute, Inc.

Abstract

A simple XML-like communications protocol is described for interactive and automated remote control of diverse instrumentation. INDI is small, easy to parse and stateless. In the INDI paradigm each Device poses all command and status functions in terms of setting and getting Properties. Each Property is a vector of one or more named members. Each Property has a current value vector; a target value vector; provides information about how it should be sequenced with respect to other Properties to accomplish one coordinated unit of observation; and provides hints as to how it might be displayed for interactive manipulation in a GUI. Clients learn the Properties of a particular Device at runtime using introspection. This decouples Client and Device implementation histories. Devices have complete authority over whether to accept commands from Clients. INDI accommodates intermediate servers, broadcasting, and connection topologies ranging from one-to-one on a single system to many-to-many between systems of different genre. The INDI protocol can be nested within other XML elements such as RTML to add constraints for automatic scheduling and execution.

Introduction


Modern astronomical instrumentation places high demand on its software infrastructure. Mounts, detectors and auxiliary equipment have many control points and configurations vary greatly across facilities and over time, creating an atmosphere of constantly changing software. Yet in the midst of this change users want consistent local and remote real-time access, freedom to use their favorite computing platforms, and ways to participate in the creation and use of clever automated scheduling techniques for efficient utilization of equipment and reduced operating costs. Keeping everything synchronized involves at least significant configuration management effort if not completely new development projects as these factors change.

This paper presents an architecture that eliminates the need for client software modifications when equipment or usage patterns change. The key feature is a communication protocol which allows a facility to describe itself in terms of devices in sufficient detail so they can be operated from a GUI for real-time command purposes or captured in a request file for later automated execution. This dynamic discovery process means that, except for the lowest level device drivers, none of the software components need to be rebuilt if and when the equipment being controlled changes. Conversely, clients may change as often as desired because they are not tied to particulars of the instrumentation. All networking and file formats use a simple subset of XML¹ which promotes freedom of choice for implementation languages and operating systems at each part of the system.

Other attempts involving XML for control purposes, such as AIML², differ from INDI in that they use XML to describe existing command and control channels. Our design uses XML directly for this purpose. Other attempts at capturing static observing requests in a file in XML format, such as the evolving RTML³, intentionally restrict themselves to the control of canonical instruments. This prevents scheduling instruments with functionality that differs from its standardized conceptual framework without changing the standard. Our design captures commands for diverse and even esoteric instrumentation with no changes in tools whatsoever.


INDI Architecture

The diagram below represents the simplest possible INDI configuration: one Client connected to one Device:



A simple INDI configuration

An INDI **Device** offers a service cast as a set of **Properties** in the INDI format. An INDI **Client** is a process which connects to an INDI Device, queries the Device for its set of control Properties and possibly send requests to change those Properties to the Device. Final authority for all instrument control rests with the Device. A Client may be a GUI which presents each command as a widget of some sort for displaying the current value of a command and possibly allowing the user to specify a new target value. A Client may be a process which never sends any changes to the Device, but monitors and logs all changes in Device state. A Client may be an automated command process that has an agenda of INDI commands and scans the agenda and decides suitable times to send new commands to the Device.



INDI with an intermediate Server

INDI Clients and Devices need not be in direct contact. The protocol is designed to accommodate arbitration and broadcasting among several Clients and Devices. For example, the diagram above shows an example of using an intermediary **Server** between INDI Clients and Devices. To each Client it appears to be a Device. To each Device it appears to be a Client. Servers can implement policies for administrative issues such as security, priority, graded access lists and other situations that arise when multiple Clients are in contact with multiple Devices.

INDI Properties

Properties are vectors of one of a small set of types. **Text** properties are collections of arbitrarily ordered characters. **Number** properties are numeric quantities and are sent with a printf-style format to recommend how a GUI should display them. **Switch** properties are always in a state of On or Off. Rules may be imposed on the behavior of the switches in a vector such as no more than one may be On at a time. **Lights** are properties that may be in one of the four states: Idle, OK, Busy or Alert. If a GUI displays this the suggested corresponding colors are gray, green, yellow and red, respectively.

Each Property has a **name** for identification purposes and a **label** for presentation purposes. Each element of the Property vector also has a name, making it in effect an associative array, and a presentation label. Changes to a Property effect all vector elements atomically.

All Property types except Lights have a **permission** attribute (lights are conceptually always read-only). Text and Number may be Read-Only, Write-Only or Read-Write; Switches may be Read-Only or Read-Write. Permission terminology is with respect to the Client but this does not bestow any true ability. Permission serves only as a hint to a Client as to whether a Device is potentially willing to allow a Property to be changed. If a Device reports a Property to be Read-Only it still must not trust the Client to comply but enforce the policy in case of rogue Clients that try to set a new value anyway. The permission hints allow Clients to treat Property values in sensible ways. For example a GUI Client may display the Property in a way that conveys whether it can be changed. For writable Properties, GUIs are encouraged to provide two fields: one that is passive and displays the last value received, and one that is interactive in which the user may type or otherwise modify a value without it being subject to spontaneous changes.

Each Property as a whole is always in one of four **states**: Idle, OK, Busy and Alert. If a GUI displays this in different colors, the colors suggested are gray, green, yellow and red, respectively. A Device is strongly encouraged to send an accompanying message whenever it informs a Client of a change of state. When a Client sends a command to change a Property, the Client shall henceforth consider the Property state to be Busy. When the Device accomplishes the associated action it sends a message that indicates the Property state has changed to, say, OK.

Each Property has a **timeout** value that specifies the worst-case time it might take to change the value to something else. The Device may report changes to the timeout value depending on current device status. Timeout values give Clients a simple ability to detect dysfunctional Devices or broken communication and also gives them a way to predict the duration of an action for scheduling purposes as discussed later.

Properties may be assembled into **groups** to suggest how Clients might organize them, for presentation purposes for example, but groups serve not functional purpose.

INDI Protocol

Each command between Client and Device specifies a **Device name** and **Property name**. The Device name serves as a logical grouping of several Properties. Property names must be unique per Device, and a Server must report unique Device names to any one Client.

The INDI protocol does not have the notion of query and response. A sender does not save context when it sends a command and wait for a specific response. A command may be sent for which a complementary command back is likely but all INDI participants must always be prepared to receive any command at any time. There is no notion of meta-errors such as illegally formatted commands, inappropriate commands or problems with the underlying communication mechanism. The proper response to all unusual or unexpected input is expressly to ignore any such problems (although some form of logging outside the scope of INDI might be judicious). With these rules the INDI protocol is small and simple; defines-away the possibility of deadlocks at the protocol level; automatically accommodates broadcasting; permits flexible and transparent routing; and eliminates the need for complex sequencing and synchronization.

When a Client first starts up it knows nothing about the Devices and Properties it will control. It begins by connecting to a Device and sending the **getProperties** command. This includes the protocol version and may include the name of a specific Device if it is known by some other means. The Device then sends back one **deftype** element for each Property it offers for control, limited to the Properties of the specified Device if included.

Note again that by sending a request for Property definitions the Client is not then waiting specifically for these definitions in reply. Nor is the Device obligated to supply these definitions in any order or particular time frame. The Client may learn of some Properties soon and perhaps others much later. The Client may also see messages for Properties about which it is as yet unaware in which case the Client silently ignores the message. Thus a Client must have the ability to dynamically expand its collection of Properties at any time and gracefully ignore unexpected information.

To inform a Device new target values for a Property, a Client sends one **newtype** element containing each member of the vector. As it does so, the Client sets its notion of the state of the Property to Busy and leaves it until told otherwise by the Device. A Client *does not* poll to see whether the current reported values happen to agree with what it last commanded and set its state back to Ok on its own. This policy allows the Device to decide how close is close enough. The Device will

eventually send a state of Ok if and when the new values for the Property have been successfully accomplished according to the Devices criteria, or send back Alert if they can not be accomplished with a message explaining why.

To inform a Client of a new set of values for a Property and their state, a Device sends one **settype** element containing each member of the vector. In the case of Properties whose values change rapidly, the Device must insure that communication of the new values are not sent so often as to saturate the connection with the Client. For example, a socket implementation might send a new value only if writing to the socket descriptor would not block the process.

A Device may send out a **message** either as part of another command or by itself. When sent alone a message may be associated with a specific Device or just to the Client in general. Messages are meant for human readers and should be sent by a Device whenever any significant event occurs or target is reached. The INDI protocol syntax provides a means for a Device to time stamp each message and is encouraged to do so to maintain consistent time across all its Clients. If the Device does not include a time stamp for some reason (such as if it is a very simple device without local time-keeping capability) the Client is encouraged to add its own time stamp.

A Device may tell a Client a given Property is no longer available by sending **delProperty**. If the command specifies only a Device without a Property, the Client must assume all the Properties for that Device, and indeed the Device itself, are no longer available.

INDI XML Syntax Specification

This section defines the exact XML syntax of each INDI command and its behavior as part of the INDI protocol. A snippet of annotated DTD is given for each command. This section ends with some examples.

The DTDs make use of the following Entities:

```
<!ENTITY % propertyState "(Idle|Ok|Busy|Alert)" >
<!ENTITY % switchState "(Off|On)" >
<!ENTITY % switchRule "(OneOfMany|AtMostOne|AnyOfMany)" >
<!ENTITY % propertyPerm "(ro|wo|rw)" >
<!ENTITY % numberValue "(#PCDATA)" >
<!ENTITY % numberFormat "(#PCDATA)" >
<!ENTITY % labelValue "(#PCDATA)" >
<!ENTITY % nameValue "(#PCDATA)" >
<!ENTITY % textValue "(#PCDATA)" >
<!ENTITY % timeValue "(#PCDATA)" >
<!ENTITY % groupTag "(#PCDATA)" >
```

All PCDATA shall use character set ISO 8651-1⁴.

The format of a numberValue shall be any one of integer, real or sexagesimal; each sexagesimal separator shall be any one of space(), colon(:) or semicolon(;); each sexagesimal component specified shall be integer or real; unspecified components shall default to 0; negative values shall be indicated with a leading hyphen(-). For example, the following are all the same numeric value: "-10:30:18", "-10 30.3" and "-10.505".

A numberFormat shall be any string that includes exactly one printf-style format specification appropriate for C-type double or one INDI style "m" to specify sexagesimal in the form "%<w>.<f>m" where

```
<w> is the total field width
<f> is the width of the fraction. valid values are:
  9 -> :mm:ss.ss
  8 -> :mm:ss.s
  6 -> :mm:ss
  5 -> :mm.m
  3 -> :mm
```

For example:

to produce	use
"-123:45"	%7.3m
" 0:01:02"	%9.6m

A timeValue shall be specified in UTC in the form YYYY-MM-DDTHH:MM:SS in accord with ISO 8601⁵.

Commands from Device to Client

In the following descriptions, permission is always with respect to the Client.

Define a property that holds one or more text elements.

```
<!ELEMENT defTextVector (defText+) >
<!ATTLIST defTextVector
  device %nameValue; #REQUIRED      name of Device
  name %nameValue; #REQUIRED        name of Property
  label %labelValue; #IMPLIED       GUI label, use name by default
  group %groupTag; #IMPLIED        Property group membership, blank by default
  state %propertyState; #REQUIRED   current state of Property
  perm %propertyPerm; #REQUIRED     ostensible Client controlability
  timeout %numberValue; #IMPLIED    worse-case time to affect, 0 default, N/A for ro
  timestamp %timeValue #IMPLIED    moment when these data were valid
  message %textValue #IMPLIED      commentary
>
```

Define one member of a text vector

```
<!ELEMENT defText %textValue >
<!ATTLIST defText
  name %nameValue; #REQUIRED      name of this text element
  label %labelValue; #IMPLIED     GUI label, use name by default
>
```

Define a property that holds one or more numeric values.

```
<!ELEMENT defNumberVector (defNumber+) >
<!ATTLIST defNumberVector
  device %nameValue; #REQUIRED      name of Device
  name %nameValue; #REQUIRED        name of Property
  label %labelValue; #IMPLIED       GUI label, use name by default
  group %groupTag; #IMPLIED        Property group membership, blank by default
  state %propertyState; #REQUIRED   current state of Property
  perm %propertyPerm; #REQUIRED     ostensible Client controlability
  timeout %numberValue; #IMPLIED    worse-case time to affect, 0 default, N/A for ro
  timestamp %timeValue #IMPLIED    moment when these data were valid
  message %textValue #IMPLIED      commentary
>
```

Define one member of a number vector

```
<!ELEMENT defNumber %numberValue >
<!ATTLIST defNumber
  name %nameValue; #REQUIRED      name of this number element
  label %labelValue; #IMPLIED     GUI label, use name by default
  format %numberFormat; #REQUIRED printf-style format for GUI display
  min %numberValue; #REQUIRED     minimal value
  max %numberValue; #REQUIRED     maximum value, ignore if min == max
  step %numberValue; #REQUIRED    allowed increments, ignore if 0
>
```

Define a collection of switches. Rule is only a hint for use by a GUI to decide a suitable presentation style. Rules are actually implemented wholly within the Device.

```
<!ELEMENT defSwitchVector (defSwitch+) >
<!ATTLIST defSwitchVector
  device %nameValue; #REQUIRED      name of Device
  name %nameValue; #REQUIRED        name of Property
  label %labelValue; #IMPLIED       GUI label, use name by default
  group %groupTag; #IMPLIED        Property group membership, blank by default
  state %propertyState; #REQUIRED   current state of Property
  perm %propertyPerm; #REQUIRED     ostensible Client controlability
  rule %switchRule; #REQUIRED       hint for GUI presentation
  timeout %numberValue; #IMPLIED    worse-case time, 0 default, N/A for ro
  timestamp %timeValue #IMPLIED    moment when these data were valid
  message %textValue #IMPLIED      commentary
>
```

Define one member of a switch vector

```
<!ELEMENT defSwitch %switchState >
<!ATTLIST defSwitch
  name %nameValue; #REQUIRED      name of this switch element
  label %labelValue; #IMPLIED     GUI label, use name by default
>
```

Define a collection of passive indicator lights.

```
<!ELEMENT defLightVector (defLight+) >
```

```

<!ATTLIST defLightVector
  device %nameValue; #REQUIRED          name of Device
  name %nameValue; #REQUIRED            name of Property
  label %labelValue; #IMPLIED           GUI label, use name by default
  group %groupTag; #IMPLIED             Property group membership, blank by default
  state %lightState; #REQUIRED          current state of Property
  timestamp %timeValue #IMPLIED         moment when these data were valid
  message %textValue #IMPLIED           commentary
>

  Define one member of a light vector
  <!ELEMENT defLight %propertyState >
  <!ATTLIST defLight
    name %nameValue; #REQUIRED          name of this light element
    label %labelValue; #IMPLIED         GUI label, use name by default
  >

Send a new set of values for a Text vector, with optional new timeout, state and message.
<!ELEMENT setTextVector (oneText+) >
<!ATTLIST setTextVector
  device %nameValue; #REQUIRED          name of Device
  name %nameValue; #REQUIRED            name of Property
  state %propertyState; #IMPLIED        state, no change if absent
  timeout %numberValue; #IMPLIED        worse-case time to affect a change
  timestamp %timeValue #IMPLIED         moment when these data were valid
  message %textValue #IMPLIED           commentary
>

Send a new set of values for a Number vector, with optional new timeout, state and message.
<!ELEMENT setNumberVector (oneNumber+) >
<!ATTLIST setNumberVector
  device %nameValue; #REQUIRED          name of Device
  name %nameValue; #REQUIRED            name of Property
  state %propertyState; #IMPLIED        state, no change if absent
  timeout %numberValue; #IMPLIED        worse-case time to affect a change
  timestamp %timeValue #IMPLIED         moment when these data were valid
  message %textValue #IMPLIED           commentary
>

Send a new set of values for a Switch vector, with optional new timeout, state and message.
<!ELEMENT setSwitchVector (oneSwitch+) >
<!ATTLIST setSwitchVector
  device %nameValue; #REQUIRED          name of Device
  name %nameValue; #REQUIRED            name of Property
  state %propertyState; #IMPLIED        state, no change if absent
  timeout %numberValue; #IMPLIED        worse-case time to affect a change
  timestamp %timeValue #IMPLIED         moment when these data were valid
  message %textValue #IMPLIED           commentary
>

Send a new set of values for a Light vector, with optional new state and message.
<!ELEMENT setLightVector (oneLight+) >
<!ATTLIST setLightVector
  device %nameValue; #REQUIRED          name of Device
  name %nameValue; #REQUIRED            name of Property
  state %propertyState; #IMPLIED        state, no change if absent
  timestamp %timeValue #IMPLIED         moment when these data were valid
  message %textValue #IMPLIED           commentary
>

Send a message associated with a device or entire system.
<!ELEMENT message >
<!ATTLIST message
  device %nameValue; #IMPLIED           considered to be site-wide if absent
  timestamp %timeValue #IMPLIED         moment when this message was generated
  message %textValue #IMPLIED           commentary
>

Delete the given property, or entire device if no property is specified.
<!ELEMENT delProperty >
<!ATTLIST delProperty
  device %nameValue; #REQUIRED          name of Device
  name %nameValue; #IMPLIED             entire device if absent
  timestamp %timeValue #IMPLIED         moment when this delete was generated
  message %textValue #IMPLIED           commentary
>

Send a message to specify state of one member of a Light vector
<!ELEMENT oneLight %propertyState >
<!ATTLIST oneLight
  name %nameValue; #REQUIRED            name of this light element
  >

```

Commands from Client to Device

Command to ask Device to define all Properties, or those for a specific Device, for which it is responsible. Must always include protocol version.

```
<!ELEMENT getProperties EMPTY >
<!ATTLIST getProperties
  version %nameValue; #REQUIRED          protocol version
  device %nameValue; #IMPLIED            properties for device, or all if absent
>
```

Commands to inform Device of new target values for a Property. After sending, the Client must set its local state for the Property to Busy, leaving it up to the Device to change it when it sees fit.

```
<!ELEMENT newTextVector (oneText+) >          send new target text values
<!ATTLIST newTextVector
  device %nameValue; #REQUIRED          name of Device
  name %nameValue; #REQUIRED           name of Property
  timestamp %timeValue #IMPLIED        moment when this message was generated
>
```

```
<!ELEMENT newNumberVector (oneNumber+) >      send new target numeric values
<!ATTLIST newNumberVector
  device %nameValue; #REQUIRED          name of Device
  name %nameValue; #REQUIRED           name of Property
  timestamp %timeValue #IMPLIED        moment when this message was generated
>
```

```
<!ELEMENT newSwitchVector (oneSwitch+) >     send new target switch states
<!ATTLIST newSwitchVector
  device %nameValue; #REQUIRED          name of Device
  name %nameValue; #REQUIRED           name of Property
  timestamp %timeValue #IMPLIED        moment when this message was generated
>
```

Elements describing a vector member value, used in both directions

One member of a Text vector

```
<!ELEMENT oneText %textValue >
<!ATTLIST oneText
  name %nameValue; #REQUIRED          name of this text element
>
```

One member of a Number vector

```
<!ELEMENT oneNumber %numberValue >
<!ATTLIST oneNumber
  name %nameValue; #REQUIRED          name of this number element
>
```

One member of a switch vector

```
<!ELEMENT oneSwitch %switchState >
<!ATTLIST oneSwitch
  name %nameValue; #REQUIRED          name of this switch element
>
```

Clients may also support *receiving* the newXXX commands above. If they choose to do so, it means to update the editable field of a writable Property in the Client GUI. Similarly, Servers may *send* all newXXX commands they receive from one Client connection to all other connected Clients. When both Client and Server support this activity, it allows Clients to mirror the commands other Clients have issued.

Example Messages from Device to Client

Define a read-write numeric field whose valid range is -100 to +100 in steps of 10, with initial value 50:

```
<defNumberVector device="OTA" name="Focus" state="Idle" perm="rw" timeout="50"
  label="Focus position, μM"
  <defNumber name="Focus" label="" format="%4.0f" min="-100" max="100" step="10">50</defNumber>
</defNumberVector>
```

Define a read-write vector with two members, one for RA and one for Dec:

```
<defNumberVector device="Mount" name="EQUATORIALJ2000_COORD" state="Idle" perm="rw" timeout="50"
  label="J2000 Equatorial Position"
  <defNumber name="RA" label="RA H:M:S" format="%11.8m" min="0" max="24">0</defNumber>
  <defNumber name="Dec" label="Dec D:M:S" format="%9.6m" min="-90" max="90">0</defNumber>
</defNumberVector>
```

Define a collection of switches for various binning settings, default setting 2:1:

```
<defSwitchVector device="Camera" name="Binning" rule="OneOfMany" state="Ok" perm="w" timeout="0"
  label="Binning">
  <defSwitch name="One" label="1:1">Off</defSwitch>
  <defSwitch name="Two" label="2:1">On </defSwitch>
  <defSwitch name="Three" label="3:1">Off</defSwitch>
  <defSwitch name="Four" label="4:1">Off</defSwitch>
</defSwitchVector>
```

Define a set of lights that indicate the security level of several alarms around the building:

```
<defLightVector device="Building" name="Security" state="Ok" label="Building Alarms">
  <defLight name="Front" label="Front door" >Ok</defLight>
  <defLight name="Back" label="Back door" >Ok</defLight>
  <defLight name="Dock" label="Loading dock">Ok</defLight>
</defLightVector>
```

Inform Client the current value of a switch has changed, assuming the switch rule was OneOfMany. Note how changes in both switches are reported, the switch change from On to Off preceding the one from Off to On and the Property state set to OK:

```
<setSwitchVector device="Camera" name="Binning" state="Ok" timestamp="2002-03-13T16:04:02"
  message="Binning changed to 1:1">
  <oneSwitch name="Two">Off</oneSwitch>
  <oneSwitch name="One">On</oneSwitch>
</setSwitchVector>
```

Send a generic system message:

```
<message timestamp="2002-03-13 12:00:00" message="It's 12 o'clock and all is well." </message>
```

Send a progress message about a device named Camera:

```
<message device="Camera" timestamp="2002-03-13 16:06:20" message="TEC is approaching target
  temperature"
</message>
```

Inform Client that the loading dock security alarm has been tripped:

```
<setLightVector device="Building" name="Security" state="Alert">
  <oneLight name="Dock">Alert</oneLight>
</setLightVector>
```

Example Messages from Client to Device

Inform Device of a new target value for a OneOfMany switch Property. Note how only the switch coming On is reported. The Client sets its local notion of the Property state to Busy. The Device implements the "one of Many" discipline, not the Client. The Device will send back the command to turn the other switch Off and set the state back to OK when the command has been accomplished. This results in completely correct behavior and relieves the Client from being expected or even trusted to do this.

```
<newSwitchVector device="Camera" name="Binning">
  <oneSwitch name="Four">On</oneSwitch>
</newSwitchVector>
```

Send new target value for RA and Dec *atomically*:

```
<newNumberVector device="Mount" name="EQUATORIALJ2000_COORD">
  <oneNumber name="RA" >10:20:30</oneNumber>
  <oneNumber name="Dec" >40:50:60</oneNumber>
</newNumberVector>
```


INDI Network Behavior

INDI is a session layer protocol. This means it *presumes* the existence of a reliable sequenced byte stream between each INDI participant. The transport mechanism for the INDI protocol may any which satisfy these requirements. It might be direct driver calls, local pipes, fifos, a UNIX socket, a TCP socket perhaps secured with SSL⁶ or tunneling through ssh⁷, or the protocol might be the payload within a framework built upon P2P⁸, JXTA⁹, Jabber¹⁰, XML-RPC¹¹ or SOAP¹². If a straight TCP/IP Socket implementation is used, the IANA¹³ has assigned INDI to tcp port 7624.

The following sections contain pseudocode that describe the required logic by which a player on an INDI network must operate. This code does not in any way attempt to address the application being accomplished by the Client or the equipment or service being operated by the Device. A key point to realize is conditions that arise that do not fit these templates, *ie*, the missing `else` clauses, are legitimately ignored although perhaps prudently logged.

Client Behavior

```

if receive <setXXX> from Device
  change record of value and/or state for the specified Property
if receive <defProperty> from Device
  if first time to see this device=
    create new Device record
  if first time to see this device+name combination
    create new Property record within given Device
if receive <delProperty> from Device
  if includes device= attribute
    if includes name= attribute
      delete record for just the given Device+name
    else
      delete all records the given Device
  else
    delete all records for all devices

if Client wants to learn all Devices and all their Properties
  send <defProperties>

if Client wants to change a Property value or state
  set State to Busy
  send <newXXX> with device, name and value

if Client wants to query a Property's target value or state
  send <getTarValue> with device and name attributes

```

Device Behavior

```

if receive <newXXX> from Client
  if element contains acceptable device, name and value
    set new target value and commence
if receive <getProperties> from Client
  if element contains recognized device
    send one <defProperty> for each name for specified device
  else if element contains no device attribute
    send one <defProperty> for each name for each device
if any Property's value changes, even as a result of a <newXXX>
  send <setXXX> specifying device+name with new value and state

```

Server Behavior

An INDI Server is really any processing steps involved with the transporting of INDI messages so long as it presents the behavior of a Client to all Devices and a Device to all Clients. In its simplest form each command the Server receives from any Device might be sent unaltered to all Clients and each command the Server receives from any Client is sent unaltered to all Devices. The efficiency of this simple Server may be improved by sending each command received from a Client to only the Device(s) known to be responsible for the `device` attribute in the command. The Server can learn these associations for itself by sending one `getProperties` to each Device just after it connects.

Automatic Scheduled Operation with INDI

So far we have discussed using INDI to control a system in real-time. The INDI self-describing paradigm can also be used to operate any system automatically. Two new XML elements are defined for the purposes of timing the actions performed. For concreteness we discuss an automated telescope application.

Scheduling involves two more components:

- a GUI that displays the Devices and Properties of a site, adds target and constraint parameters and saves the entire description to a request file;
- an INDI client serving as a scheduling daemon that is in an infinite loop: scanning all requests to select the next observation and communicating with the site Devices to accomplish the work.

The schedule preparation GUI could look very much like the real-time GUI, perhaps even be the same program. It certainly shows each Device and Property in much the same way. But it also contains a second section which allows the user to define a targeting scheme, timing requirements and to place constraints on the observation.

Observation timing might be specified as a particular Date and Time, or JD, with an acceptable error margin; be specified as a cyclical event with reference, period and phase timing window specifications; or might not be specified at all and be left totally up to the scheduler. Whichever way the time is decided, it serves to define the reference start time of the observation.

Other constraints might be a required such as maximum seeing spread or an acceptable range of Moon phase. In the former case, the scheduler could know to ask the Device for the current value of a read-only property which reports the seeing from, say, a DIMM¹⁴ device. In the latter case, the daemon consults its own lunar ephemeris and knowledge of the location of the site.

To capture the timing of events comprising one observation, two more elements are added to the XML INDI protocol. An INDI *newtype* command can be wrapped in a **by** element with an attribute **t** that specifies with respect to the start of the observation the number of seconds the Property should achieve status OK. At runtime the Scheduler can determine the time it takes to achieve a given target value using the timeout parameter for the Property.

An INDI *newtype* command can also be wrapped in an **at** element, also with an attribute **t**. This tells the scheduler Client the exact moment when the *newtype* should be issued. It can be used with any Property but makes most sense when used with ones whose timeout value is 0, for example a *switch* that takes no time to accomplish such as a camera shutter control.

The role of the Scheduler process is to keep the telescope busy. Whenever the telescope is not busy, the scheduler scans and ranks all observing requests still to be performed. The ranking includes local circumstances such as hour angle and imminent sun rise and set. Constraints are determined to be met by asking the Devices to report the current property values and status. When a request is finally chosen, it is executed by simply issuing the INDI property commands *exactly as they appear in the file*, timed according to their **at** or **by** attributes.

Operational specifications use the Devices and Properties for the observatory directly to record which commands are to be issued for the observation. The scheduler will cancel the observation if any of these return other than *Busy* or *OK* status values. For read-only properties, the scheduler requires that the Devices return values that match those in the observation request before choosing that observation to execute.

The following is an example of one observation request. Each of the Constraints were specified by the user in the Schedule Builder GUI. The Operations were compiled using the exact Device and Property names returned to the Builder by the Devices for which the schedule was built. Note the ability to specify different property values to occur at different times; in this case, to implement a multiple exposure to determine the direction of any trailed objects in the image.

```

<INDIObservation>
  <moon>
    <minSeparation>20</minSeparation>
    <maxPercentLit>10</maxPercentLit>
  </moon>
  <timing method="phase">
    <phaseTiming>
      <phaseZeroTime>2345678.901</phaseZeroTime>
      <phasePeriod>3.45</phasePeriod>
      <phaseMin>.3</phaseMin>
      <phaseMax>.4</phaseMax>
    </phaseTiming>
  </timing>
  <repeat>
    <count>10</count>
    <minSep>1</minSep>
  </repeat>
  <INDI>
    <by t="0">
      <newNumberVector device="Monster Scope" name="EQUATORIALJ2000_COORD">
        <oneNumber name="RA">10:20:30</oneNumber>
        <oneNumber name="Dec">-4:5:6</oneNumber>
      </newNumberVector>
      <newTextVector device="OTA" name="Big-O Filters">
        <oneText name="setting">Red</oneText>
      </newTextVector>
    </by>
    <at t="0">
      <newSwitchVector device="Wonder Cam" name="Shutter">
        <oneSwitch name="open">On</oneSwitch>
      </newSwitchVector>
    </at>
    <at t="10">
      <newSwitchVector device="Wonder Cam" name="Shutter">
        <oneSwitch name="open">Off</oneSwitch>
      </newSwitchVector>
    </at>
    <at t="20">
      <newSwitchVector device="Wonder Cam" name="Shutter">
        <oneSwitch name="open">On</oneSwitch>
      </newSwitchVector>
    </at>
    <at t="60">
      <newSwitchVector device="Wonder Cam" name="Shutter">
        <oneSwitch name="open">Off</oneSwitch>
      </newSwitchVector>
    </at>
  </INDI>
  <Log>
  </Log>
</INDIObservation>

```

As the Scheduler daemon executes each command for this request, the Log portion is filled in. Each entry in the Log is timestamped in format ISO 8601. This forms a permanent record of all actions and results in behalf of this observation.

Summary

INDI Clients know nothing about the meaning of the Devices and Properties with which they deal and Devices know nothing of how or why Clients are using their services. Devices present a standard protocol to all potential Clients, so they need only be written one time (for a given operating system) and thus are defined by and move with their equipment and not with their control programs. Clients control any Devices. A real-time GUI Client just displays Properties and lets the user change some of them. A schedule-creation GUI displays the Properties and saves their target values to an XML file containing *newtype* elements along with additional tags to define constraints and timing. A scheduler daemon reads the file and copies the tags directly to the devices at the appropriate time.

Thus, GUIs, file formats and control daemons remain completely unchanged and drivers are only written once to be accessed from anywhere making INDI truly an *Instrument Neutral Distributed Interface*.

Appendix I. Standardized Properties

Since INDI itself imposes no meaning on Properties, the key to interoperability is to agree on names for Properties that occur often in a discipline. We propose the following as a rudimentary beginning for the Astronomical application of INDI:

Property name	Type	Member name	Description
EQUATORIAL_COORD	Number	RA DEC	Equatorial astrometric J2000 coordinate J2000 RA, hours J2000 Dec, degrees +N
EQUATORIAL_EOD_COORD	Number	RA DEC	Equatorial apparent coordinate at Equinox of Date EOD RA, hours EOD Dec, degrees +N
GEOGRAPHIC_COORD	Number	LATITUDE LONGITUDE	Earth geodetic coordinate Latitude, degrees +N Longitude, degrees +E
HORIZONTAL_COORD	Number	ALTITUDE AZIMUTH	Degrees above horizon Degrees E of N
ABORT_MOTION	Switch	ABORT	Stop rapidly but gracefully
ON_COORD_SET	Switch	SLEW TRACK SYNC	Action device takes when sent any *_COORD property Slew to a coordinate and stop Slew to a coordinate and track Accept current coordinate as correct
POWER	Switch	ON OFF	Turn on power supply to this device Turn off power supply to this device

Resources

-
- 1 <http://www.w3.org/TR/2000/REC-xml-20001006>
 - 2 <http://pioneer.gsfc.nasa.gov/public/aiml>
 - 3 <http://alpha.uni-sw.gwdg.de/~hessman/RTML/RTML3>
 - 4 <http://www.htmlhelp.com/reference/charset/latin1.gif>
 - 5 <http://www.cl.cam.ac.uk/~mgk25/iso-time.html>
 - 6 <http://home.netscape.com/eng/ssl3/draft302.txt>, <http://www.openssl.org>
 - 7 <http://www.openssh.com>
 - 8 <http://www.openp2p.com>
 - 9 <http://www.jxta.org>
 - 10 <http://www.jabber.org>
 - 11 <http://xml-rpc.com>
 - 12 <http://www.w3.org/TR/SOAP>
 - 13 <http://www.iana.net>
 - 14 <http://www.astro.washington.edu/rest/dimm>